

Nickel–Titanium Rotary Instruments: An *In Vitro* Comparison

Torsional Resistance of Two Heat-treated Reciprocating Files

Shilpa Bhandi¹, Marco Seracchiani², Orlando Donfrancesco³, Rodolfo Reda⁴, Alessandro Mazzoni⁵, Stefania Nottola⁶, Giuseppe Familiari⁷, Luca Testarelli⁸, Gianluca Gambarini⁹

ABSTRACT

Aim: The present study aims to evaluate the difference in torsional resistance of two reciprocating nickel–titanium (NiTi) rotary files: WaveOne Gold and EdgeOne Fire.

Materials and methods: A total of 40 nickel–titanium rotary instruments ($n = 40$): 20 WaveOne Gold Small (WOGS) and 20 EdgeOne Fire Small (EOFS) were divided into two groups. Each instrument was tested using a torsional resistance device already validated in previous studies to evaluate and compare torsional resistance. The static torsional test was implemented by blocking each instrument at 3 mm from the tip and rotating it until fracture with a reciprocating motion. Torque to fracture (TtF) and fragment length (FL) were measured and statistically analyzed.

Results: Statistical analysis of TtF found significant differences between the two groups ($p < 0.05$). The EOFS showed higher TtF if compared to WOGS, with a mean value and a standard deviation of 3.05 ± 0.07 (N cm) against 2.97 ± 0.08 (N cm). Data for FL showed no significant differences ($p > 0.05$) between the two groups.

Conclusion: According to the results of this study, it is reasonable to assert that EOFS instruments showed a higher torsional resistance if compared to the WOGS.

Clinical significance: As evidenced by this study, EOFS should be considered as a safer solution, in terms of torsional resistance, if compared to WOGS, reducing the risk of intracanal separation due to excessive torsional load.

Keywords: EdgeOne Fire, Endodontics, Nickel–titanium file, Torsional resistance, WaveOne Gold.

The Journal of Contemporary Dental Practice (2021): 10.5005/jp-journals-10024-3081

INTRODUCTION

The introduction of nickel–titanium rotary (NTR) instruments in endodontics has significantly reduced the operative time of intracanal instrumentation.¹ Moreover, NTR files improved the shaping and cleaning ability of the root canal during clinical procedures.² Despite these superior qualities, the prolonged use of NTR instruments can lead to fracture during the root canal treatment (RCT).³ According to Sattapan et al., two main patterns of fracture can occur using nickel–titanium (NiTi) files: torsional and flexural.^{4,5} These two patterns of failure could be considered as the major drawback during RCTs. Flexural stresses could result in the formation of microcracks on the instrument surface and, after repetitive compression and tension, the instrument can be fractured due to cyclic fatigue.⁴ File failure due to torsional stress occurs when the tip or part of the file remains blocked between the canal walls while it continues to rotate.^{4,6} For this reason, in recent literature many studies focused their attention on NiTi torsional behavior.^{7,8} Recent studies investigated the role of two prevailing factors affecting torsional resistance, that is, heat treatments and the metal mass of the instrument.^{7–9} According to the most recent literature, static torsional resistance tests have been evaluated as described by International Organization for Standardization (ISO) 3630-1. According to this, to evaluate torsional resistance, a stainless steel (SS) instrument is rotated at 2 rpm with its tip blocked at 3 mm until fracture and the torque generated is calculated by a torque meter. However, the ISO 3630-1 evaluation does not consider that during clinical practice, NTR instruments are rotated at a higher speed if compared to SS files. According to the authors, the above-mentioned ISO 3630-1 should be changed for a methodology that takes into account the clinical application of NTR files, since it

¹Department of Restorative Dental Sciences, Jazan University, Jazan, Saudi Arabia

^{2,4,5,8,9}Department of Oral and Maxillo Facial Sciences, Sapienza University of Rome, Rome, Italy

^{3,6,7}Department of Anatomy, Histology, Forensic Medicine and Orthopaedics, Sapienza University of Rome, Rome, Italy

Corresponding Author: Rodolfo Reda, Department of Oral and Maxillo Facial Sciences, Sapienza University of Rome, Rome, Italy, Phone: +39 3287353196, e-mail: rodolforeda17@gmail.com

How to cite this article: Bhandi S, Seracchiani M, Donfrancesco O, et al. Nickel–Titanium Rotary Instruments: An In Vitro Comparison. *J Contemp Dent Pract* 2021;x(x):xx–xx.

Source of support: Nil

Conflict of interest: None

has been demonstrated that rotational speed does not affect the torsional resistance of NTR instruments.¹⁰

In the last decades, new kinematics were introduced to minimize the influence of flexural and torsional stresses during instrumentation.¹¹ Reciprocating motion was introduced to improve resistance to fracture since it reduces cyclic and torsional fatigue during root canal treatment.^{12,13} Reciprocating motion consists of development of a cutting angle and a releasing angle during the instrumentation. The subsequent alternating changes in the direction of rotation could reduce the cyclic fatigue of the NTR files if compared to rotating motion.¹⁴ WaveOne Gold (WOG) (Dentsply Maillefer, Ballaigues, Switzerland) is a recent single-file reciprocating instrument that replaced WaveOne (Dentsply Maillefer, Ballaigues, Switzerland). Concerning the cross-sectional design, the WOG is

characterized by an uncentred parallelogram with two cutting edges along the instrument surface.¹⁵ Moreover, WOG presents the Gold Wire (Dentsply Maillefer, Ballaigues, Switzerland) heat treatment, which according to Zupanc et al. is characterized by a martensitic status with varying amounts of austenite and R-phase.¹⁶ These features enhance the instrument with a controlled memory effect and flexibility.¹⁶

EdgeOne Fire (EOF) (EdgeEndo, Albuquerque, NM, USA) is a new instrument with the same cross-section of the WOG. This instrument presents a proprietary heat treatment called FireWire™ (EdgeEndo, Albuquerque, NM, USA). According to the manufacturer's declaration, this heating process provides high flexibility and superior resistance to cyclic fatigue tests. This instrument is developed to be used in reciprocating motion, with the same handpiece settings adopted for WaveOne Gold.¹⁷ To date, no studies in the literature have evaluated the torsional behavior of these NTR instruments yet. The present study aimed to evaluate the differences in terms of torsional resistance at fracture between WaveOne Gold and EdgeOne Fire with a previously validated torsional resistance testing device.

MATERIALS AND METHODS

The study has been performed at the Department of Oral and Maxillo Facial Sciences of Sapienza University of Rome (Rome, Italy).

After calculating the sample size, and in agreement with the recently published literature on this subject, 40 NTR files ($n = 40$) were used in this study, divided into two groups using a simple randomization technique: 20 WaveOne Gold Small (Group A) and 20 EdgeOne Fire Small (Group B). All the tested instruments were evaluated with $\times 20$ magnification (SMZ 745T; Nikon, Tokyo, Japan) before undergoing the tests. None of the analyzed NTR instruments were discarded due to manufacturing defects, such as microcracks and dented cutting edge.

The torsional test was performed using a torsional resistance testing device that was already validated in the literature.^{18,19} A torque recording endodontic motor (KaVo, Biberac, Germany) appropriately modified to perform reciprocating motion, was used in this study. This endodontic motor could register the variations of the torque generated by the instrument during the intracanal instrumentation every 0.1 seconds. According to ISO 3630-1, the tip of the file was blocked at 3 mm from the tip, stuck in a vise (Fig. 1).



Fig. 1: Torsional static test device




Fig. 2: Microscope analysis of EdgeOne Fire Small. Torsional fracture is evidenced by the round-shaped circumferential lines³²




Fig. 3: Microscope analysis of WaveOne Gold Small. Torsional fracture is evidenced by the round-shaped circumferential lines³²

Using the same settings recommended for WaveOne Gold, each file was rotated using reciprocating motion, until fracture occurred. The torque limit was set at 5.5 N cm. Torque to fracture (TtF) was recorded and tips fragment length (FL) of the instrument was measured using a digital caliper. Moreover, the separated instruments of both groups were analyzed with a Zeiss Axio Microscope (Carl Zeiss MicroImaging, Göttingen, Germany) to visualize fracture patterns (Figs 2 and 3).

Data were collected and analyzed using the SPSS 17.0 software (SPSS Incorporated, Chicago, IL, USA). Moreover, the mean value and standard deviation were calculated and the student's *t* test with the significance level set at $p < 0.05$ was performed.

RESULTS

Table 1 shows the mean values for FL for both groups. Statistical analysis showed no significant differences ($p > 0.05$) between the two instruments.

Table 2 shows the mean values of TtF for both groups. Group B showed higher TtF if compared to Group A.

Table 1: Mean values for fragment length (FL) for both groups A and B in mm ($p > 0.05$)

Mean values for FL in mm (mean \pm SD)	
Group A	Group B
2.97 \pm 0.08	3.05 \pm 0.07

Table 2: Mean values for torque to fracture (TtF) for both groups A and B in N cm ($p < 0.05$)

Mean values for TtF in N cm (mean \pm SD)	
Group A	Group B
0.27 \pm 0.01	0.35 \pm 0.04

Statistical analysis found significant differences between the two groups ($p < 0.05$). According to the results, EdgeOne Fire showed higher TtF if compared to WaveOne Gold.

DISCUSSION

Instrument failure during RCTs remains a serious drawback for the clinician.^{20,21} Root canal geometry may be considered as a leading factor influencing the stress applied to the instrument and the high fracture incidence in curved canals is a consequence of this.³ As described by Parashos and Messer, another important factor in determining fatigue resistance of NTR instruments is the number of uses and stress accumulated during previous RCTs.^{3,22} Furthermore, the first shaping file of a sequence can be subjected to a higher torsional load during instrumentation.²³ A recently published study showed that the non-use of lubricants during intracanal instrumentation could increase the instrument fracture due to torsional stress.²⁴ For the above-mentioned reasons, the reciprocating motion has been proposed to decrease the frequency of taper locks.²⁵ The NTR instruments presented in these studies, WaveOne Gold and EdgeOne Fire, have been designed to be used with reciprocating motion. As described above, the alternating motion gives superior flexural resistance to the instruments during clinical usage.²⁵

NTR instruments fracture usually without any visual warnings and with only microscopical permanent deformations. In literature, many studies investigated the main features that could influence the torsional resistance of instruments: numerous variables, like presence of flexural stress, instrument manufacturing processes, taper, size, pitch, shaft length, cross-sectional design, and metal mass distributed along the instruments could influence the clinical performance of an endodontic file and its resistance to both cyclic and torsional stresses.^{3,26–30}

Since multiple parameters can affect the torsional strength, during *in vitro* tests, it is necessary to consider only one feature to avoid bias. Indeed, the major problem in carrying out this type of research is the standardization of each of the above-mentioned factors. For these reasons, we selected WaveOne Gold and EdgeOne Fire, which present the same tip, taper, cross-sectional design, and pitch, and they differ only for the heat treatments provided by the manufacturers. Moreover, the manufacturers suggest the same clinical motion for the two files, with the same cutting and releasing angle.

It has been demonstrated that heat-treated NiTi instruments have a better performance regarding the angular rotation and torsional resistance if compared to conventional similar NiTi files. As described by Tabassum et al., the extreme ductility provided

by some heat treatments could explain the superior flexibility and therefore the higher resistance to failure of such type of instruments.³¹

As showed in Table 2 fragment length (FL) mean values were calculated and no significant differences were found ($p > 0.05$). This result demonstrates that there is no difference in the tip length of the files blocked for the test and, therefore, no differences in the portion of the instrument subjected to the stresses.

Results for TtF showed significant differences ($p < 0.05$) for both tests. These results could be explained by the prevalent difference between the two tested instruments. WaveOne Gold presents an innovative heat treatment, the Gold Wire. To date, many studies investigated the role of Gold Wire alloy in torsional and flexural resistance. Gold treated instruments represent new technology in the field of endodontic instruments resistance.¹⁶ Therefore, in this study, the EdgeOne Fire instruments resulted to be more resistant to torsional stress when compared to identical instruments made with Gold Wire treatment. Even if the number of the instruments used in the present study is meager, the difference between the two tested instruments could be due to the innovative three-dimensional aspect of the crystalline matrix of the FireWire™ that confers to the alloy not only higher flexibility, as shown in previously published studies, but also superior resistance to torsional stress.¹⁷

CONCLUSIONS

Since we used the same motion and similar-in-shape instruments, differences could be related to the different heat treatments provided during the manufacturing process. Within the limitation of the present study, it can be assessed that the FireWire™ treatment could provide more torsional resistance to the instruments if compared to the Gold Wire. Further studies are needed to better evaluate the role of thermal treatments during torsional resistance tests.

CLINICAL SIGNIFICANCE

As evidenced by the results of this study, the possibility of reaching a higher torsional resistance by performing proper heat treatments during the manufacturing process should be considered a very promising solution that could improve endodontic practice, reducing the clinical risk of intracanal separation and its consequent complication in reaching a proper RCT.

REFERENCES

- Glossen CR, Haller RH, Dove SB, et al. A comparison of root canal preparations using Ni-Ti hand, Ni-Ti engine-driven, and K-Flex endodontic instruments. *J Endod* 1995;21(3):146–151. DOI: 10.1016/s0099-2399(06)80441-3.
- Htun PH, Ebihara A, Maki K, et al. Comparison of torque, force generation and canal shaping ability between manual and nickel-titanium glide path instruments in rotary and optimum glide path motion. *Odontology* 2020;108(2):188–193. DOI: 10.1007/s10266-019-00455-1.
- Parashos P, Messer HH. Rotary NiTi instrument fracture and its consequences. *J Endod* 2006;32(11):1031–1043. DOI: 10.1016/j.joen.2006.06.008.
- Sattapan B, Nervo GJ, Palamara JE, et al. Defects in rotary nickel-titanium files after clinical use. *J Endod* 2000;26(3):161–165. DOI: 10.1097/00004770-200003000-00008.
- Seracchiani M, Miccoli G, Reda R, et al. A Comprehensive *in vitro* comparison of mechanical properties of two rotary endodontic

- instruments. *World J Dent* 2020;11(3):185–188. DOI: 10.5005/jp-journals-10015-1729.
6. Alapati SB, Brantley WA, Svec TA, et al. SEM observations of nickel-titanium rotary endodontic instruments that fractured during clinical use. *J Endod* 2005;31(1):40–43. DOI: 10.1097/01.don.0000132301.87637.4a.
 7. Gambarini G, Miccoli G, Di Nardo D, et al. Torsional resistance of two new heat treated nickel titanium rotary instruments: an in vitro evaluation. *Pesqui Bras Odontopediatria Clin Integr* 2020;20:0053.
 8. Gambarini G, Cicconetti A, Di Nardo D, et al. Influence of different heat treatments on torsional and cyclic fatigue resistance of nickel-titanium rotary files: a comparative study. *Appl Sci (Switzerland)* 2020;10(16):5604. DOI: 10.3390/app10165604.
 9. Di Nardo D, Gambarini G, Seracchiani M, et al. Influence of different cross-section on cyclic fatigue resistance of two nickel-titanium rotary instruments with same heat treatment: an in vitro study. *Saudi Endod J* 2020;10(3):221–225. DOI: 10.4103/sej.sej_124_19.
 10. Ha JH, Kwak SW, Kim SK, et al. Effect from rotational speed on torsional resistance of the nickel-titanium instruments. *J Endod* 2017;43(3):443–446. DOI: 10.1016/j.joen.2016.10.032.
 11. Ferreira F, Adeodato C, Barbosa I, et al. Movement kinematics and cyclic fatigue of NiTi rotary instruments: a systematic review. *Int Endod J* 2017;50(2):143–152. DOI: 10.1111/iej.12613.
 12. Kiefner P, Ban M, De-Deus G. Is the reciprocating movement per se able to improve the cyclic fatigue resistance of instruments? *Int Endod J* 2014;47(5):430–436. DOI: 10.1111/iej.12166.
 13. Tokita D, Ebihara A, Miyara K, et al. Dynamic torsional and cyclic fracture behavior of profile rotary instruments at continuous or reciprocating rotation as visualized with high-speed digital video imaging. *J Endod* 2017;43(8):1337–1342. DOI: 10.1016/j.joen.2017.03.024.
 14. Gambarini G, Rubini AG, Al Sudani D, et al. Influence of different angles of reciprocation on the cyclic fatigue of nickel-titanium endodontic instruments. *J Endod* 2012;38(10):1408–1411. DOI: 10.1016/j.joen.2012.05.019.
 15. Kwak SW, Lee CJ, Kim SK, et al. Comparison of screw-in forces during movement of endodontic files with different geometries, alloys, and kinetics. *Materials (Basel)* 2019;12(9):1506. DOI: 10.3390/ma12091506.
 16. Zupanc J, Vahdat-Pajouh N, Schäfer E. New thermomechanically treated NiTi alloys—a review. *Int Endod J* 2018;51(10):1088–1103. DOI: 10.1111/iej.12924.
 17. Gambarini G, Galli M, Di Nardo D, et al. Differences in cyclic fatigue lifespan between two different heat treated NiTi endodontic rotary instruments: WaveOne Gold vs EdgeOne Fire. *J Clin Exp Dent* 2019;11(7):e609–e613. DOI: 10.4317/jced.55839.
 18. Di Nardo D, Seracchiani M, Mazzoni A, et al. Torque range, a new parameter to evaluate new and used instrument safety. *Appl Sci* 2020;10(10):3418. DOI: 10.3390/app10103418.
 19. Gambarini G, Miccoli M, D'Angelo M, et al. The relevance of operative torque and torsional resistance of nickel-titanium rotary instruments: a preliminary clinical investigation. *Saudi Endod J* 2020;10:260–264. DOI: 10.4103/sej.sej_157_19.
 20. Madarati AA. Factors influencing incidents of complications while using nickel-titanium rotary instruments for root canal treatment. *BMC Oral Health* 2019;19(1):241. DOI: 10.1186/s12903-019-0938-7.
 21. Miccoli G, Seracchiani M, Zanza A, et al. Possible complications of endodontic treatments. *J Contemp Dent Pract* 2020;21(5):473–474.
 22. Miccoli G, Seracchiani M, Del Giudice A, et al. Fatigue resistance of two nickel-titanium rotary instruments before and after ex vivo root canal treatment. *J Contemp Dent Pract* 2020;21(7):728–732.
 23. Kwak SW, Ha JH, Cheung GS, et al. Effect of the glide path establishment on the torque generation to the files during instrumentation: an in vitro measurement. *J Endod* 2018;44(3):496–500. DOI: 10.1016/j.joen.2017.09.016.
 24. Mazzoni A, Pacifici A, Zanza A, et al. Assessment of real-time operative torque during nickel-titanium instrumentation with different lubricants. *Appl Sci* 2020;10(18):6201. DOI: 10.3390/app10186201.
 25. Kimura S, Ebihara A, Maki K, et al. Effect of optimum torque reverse motion on torque and force generation during root canal instrumentation with crown-down and single-length techniques. *J Endod* 2020;46(2):232–237. DOI: 10.1016/j.joen.2019.11.007.
 26. Baek SH, Lee CJ, Versluis A, et al. Comparison of torsional stiffness of nickel-titanium rotary files with different geometric characteristics. *J Endod* 2011;37(9):1283–1286. DOI: 10.1016/j.joen.2011.05.032.
 27. Alcalde MP, Tanomaru-Filho M, Bramante CM, et al. Cyclic and torsional fatigue resistance of reciprocating single files manufactured by different nickel-titanium alloys. *J Endod* 2017;43(7):1186–1191. DOI: 10.1016/j.joen.2017.03.008.
 28. Gambarini G, Miccoli G, Seracchiani M, et al. Role of the flat-designed surface in improving the cyclic fatigue resistance of endodontic NiTi rotary instruments. *Materials (Basel)* 2019;12(16):2523. DOI: 10.3390/ma12162523.
 29. Gambarini G, Seracchiani M, Zanza A, et al. Influence of shaft length on torsional behavior of endodontic nickel-titanium instruments. *Odontology* 2020. DOI: 10.1007/s10266-020-00572-2. Epub ahead of print. PMID: 33245455.
 30. Seracchiani M, Miccoli G, Di Nardo D, et al. Effect of flexural stress on torsional resistance of NiTi instruments. *J Endod* 2021;47(3):472–476. DOI: 10.1016/j.joen.2020.10.011. Epub 2020 Oct 21. PMID: 33096192.
 31. Tabassum S, Zafar K, Umer F. Nickel-titanium rotary file systems: what's new? *Eur Endod J* 2019;4(3):111–117. DOI: 10.14744/eej.2019.80664.
 32. Zupancich A, Cristiani E. Functional analysis of sandstone ground stone tools: arguments for a qualitative and quantitative synergetic approach. *Sci Rep* 2020;10(1):15740. DOI: 10.1038/s41598-020-72276-0.